

Troubleshooting

Problem	Potential Cause(s)	Recommended Solution
No FISH signals are detected in the microscope.	<ul style="list-style-type: none">Reflected light shutter closed / stop slider in light path.Fluorescent lamp is switched off.Wrong fluorescence filter is in light path.Objective is out of position.Phototube is in camera position.	<ul style="list-style-type: none">Open shutter / move stop slider out of the light path.Switch on fluorescent lamp.Move correct filter into light path.Swing objective into light path.Direct light path to eyepieces.
Hybridization signals become weak after a while.	<ul style="list-style-type: none">Immersion oil soaked in-between slide and coverslip.	<ul style="list-style-type: none">Replace coverslip and DAPI/antifade. Use 24 x 32 mm² coverslip even if only a small region is hybridized.
Diffuse signals.	<ul style="list-style-type: none">Preparation is not adequately illuminated.Focus plane cannot be adjusted properly.Antifade layer is too thick for focusing.	<ul style="list-style-type: none">Check optical pathway of microscope. Adjust the UV light properly. Check the lifetime of the UV lamp.Use enough immersion oil. Do not mix different immersion oils. Use immersion oil suitable for fluorescence.Do not use too much DAPI/antifade. 10 µl per slide (24 x 32 mm² coverslip) are sufficient.
Weak signals.	<ul style="list-style-type: none">Chromosome slide preparation is too old.Denaturation of chromosomes is not adequate.A multi bandpass filter is used for viewing.	<ul style="list-style-type: none">Slides should not be older than two weeks.Aging, baking or further fixation may inhibit the hybridization and is not recommended.Increase denaturation temperature up to 80°CUse a dedicated single bandpass filter.
Weak aqua or green signals or high diffuse background in green color channel.	<ul style="list-style-type: none">DAPI intensity is too high resulting in crosstalk to AQUA filter or GREEN filter.pH value of washing solutions is too low.	<ul style="list-style-type: none">Use DAPI/antifade of low concentration.Ensure that pH value is between 7.0 and 7.5 of solutions. Some green fluorophores are very sensible to pH below 7.
High unspecific background	<ul style="list-style-type: none">Remaining cytoplasmic proteins of the cells may impair the hybridization.	<ul style="list-style-type: none">Pretreat slides with Pepsin.
If the recommended measures do not solve the problem, or your problem is not listed, please contact MetaSystems Probes.		









Customer Support

Please contact MetaSystems Probes GmbH (contact details see below) or our authorized distributor in your country. MetaSystems Probes disclaims any proprietary interest in the marks and names of others.



MetaSystems Probes GmbH
1. Industriestraße 7
68804 Altlussheim
Germany
Tel.: +49 (0)6205 292760
Fax: +49 (0)6205 2927629
email: info@metasystems-probes.com
URL: www.metasystems-probes.com

Symbols Used

Symbol	Description
	This symbol marks a product as an “In Vitro Diagnostic Medical Device”.
	All warnings are marked by warning triangle with exclamation mark. Depending on their character they are supplemented with the words ATTENTION or CAUTION.
	Manufacturer
	Reference number
	No of tests
	Lot number
	Expiry date
	Temperature limitation for storage. Lower and upper limits are indicated.



FORMAMIDE
Danger. May damage the unborn child. Suspected of causing cancer. May cause damage to organs through prolonged/repeated exposure. Obtain special instructions before use. Do not breathe vapours. Wear protective gloves/protective clothing. IF exposed or concerned: Get medical advice.

Gefahr. Kann das Kind im Mutterleib schädigen. Kann vermutlich Krebs erzeugen. Kann die Organe schädigen bei längerer/wiederholter Exposition. Vor Gebrauch besondere Anweisungen einholen. Dampf nicht einatmen. Schutzhandschuhe/Schutzkleidung tragen. Bei Exposition oder Verdacht: Ärztlichen Rat einholen.

Danger. Peut nuire au fœtus. Susceptible de provoquer le cancer. Risque présumé d'effets graves pour les organes à la suite d'expositions répétées ou d'une exposition prolongée. Se procurer les instructions avant utilisation. Ne pas respirer les vapeurs. Porter des gants de protection/ des vêtements de protection. EN CAS d'exposition prouvée ou suspectée: consulter un médecin.

MetaSystems Probes GmbH
1. Industriestr. 7, 68804 Altlussheim,
Germany, Tel.: +49 (0)6205 292760



XL ROS1-GOPC BA
Break Apart Probe
100µl ( 10)
REF: D-6029-100-OG 10



XXXXX



YYYY-MM-DD

-25°C  -15°C



XCyting Locus-Specific Probes


For Professional Use Only

Further information available at www.metasystems-probes.com

Product	Label	Order No.	Pack Size
XL ROS1-GOPC BA	orange/green	D-6029-100-OG	100µl

The XL ROS1-GOPC BA probe is designed as a break apart probe. Its green labeled part hybridizes proximal to the ROS1 gene at 6q22 and the orange labeled probe hybridizes to the distal region of ROS1. This probe is also suited to analyze the ROS1-GOPC (formerly called FIG-ROS) fusion, caused by an interstitial deletion of 240 kb in size. This probe is intended for methanol/acetic-acid fixed cells and tissue sections.

Probe Diagram:



Materials Provided





100µl of XL ROS1-GOPC BA, the probe mix is dissolved in hybridization solution and ready to use.

Intended Use

DNA FISH probes are intended for fluorescence in-situ hybridization (FISH) for the analysis of chromosomal aberrations on fixed cells from human tissue suitable for cytogenetic investigation. Hybridized to metaphase and/or interphase nuclei FISH probes allow the analysis of chromosome structure or copy number variations to detect acquired genetic alterations according to the Global Medical Device Nomenclature (GMDN) CT929. FISH analysis is used as an adjunct test to other diagnostic investigations and not to be used as sole base for diagnosis or therapy decisions.

Safety Instructions

All probes produced by MetaSystems Probes are for professional use only and should be used by qualified and trained personnel only. In order to ensure safe operation and reproducible results please observe the safety notices and caution signs below.

	CAUTION: Formamide is toxic and a potential teratogen! MetaSystems probes contain formamide. Formamide is toxic and a teratogen. May cause harm to the unborn child. Do not breathe vapours; avoid skin contact! Wear gloves and a lab coat. In case of contact with skin or eyes, wash immediately with water.
	CAUTION: Hot water bath and hot plates! For denaturation and hybridization hot water baths and hot plates are used with temperatures of >37°C. Be careful not to get in direct contact with hot surfaces or liquids. Wear gloves and a lab coat. In case of contact with skin, cool immediately with cold water.
	ATTENTION: Good Laboratory Practice! Use in accordance with the principles of good laboratory practice.
	ATTENTION: Waste Disposal! All hazardous materials should be disposed of according to local/ national regulation for hazardous waste disposal.

Storage and Handling

Probes should be stored in the dark at -20°C (±5°C). Probe performance has been shown to be unaffected for up to 20 freeze-thaw cycles.

Shipping

MetaSystems' DNA probes are shipped at room temperature.

Equipment Necessary but not Supplied

<ul style="list-style-type: none">• Water bath with accurate temperature control• Variable micro-pipettes with volumes ranging from 1 µl to 1 ml, calibrated• Thermometer• pH meter, calibrated	<ul style="list-style-type: none">• Hotplate 75°C (±1°C), with a solid plate and accurate temperature control up to 80°C• Freezer -20°C (±5°C)• Humidified chamber 37°C (±1°C)• Forceps• Gloves• Microcentrifuge	<ul style="list-style-type: none">• Fluorescence microscope with suitable filters (see below)• Immersion oil, recommended by the microscope manufacturer (fluorescence grade)• Imaging System, e. g. Isis (MetaSystems)• Coverslips (glass): 22 x 22 mm² and 24 x 32 mm²• Rubber Cement• DAPI/antifade
--	---	---

Fluorescence Microscope Recommendation

- Fluorescence Illumination: Metal halide fluorescence illumination systems or conventional 100 watt mercury lamp illuminators
- Objectives suitable for epi-fluorescent illumination.
- Fluorescence Filters: For viewing/counting use a MetaSystems triple or quad bandpass filter set or appropriate single bandpass filter. For capturing images use suitable single bandpass filters for the respective fluorochromes. Please inquire.

Fluorochrome Specification

Label	Absorption max.	Emission max.
Blue (aqua)	426 nm	480 nm
Green	505 nm	530 nm
Orange	552 nm	576 nm

Sample Preparation

General Comments

- MetaSystems probes are designed for use on cytogenetic samples which are fixed in 3:1 methanol/acetic-acid and should be prepared according to the laboratory or institution guidelines.
- Prepare specimen according to standard cytogenetic procedures.

Stability of Hybridized Slides

- Hybridized FISH slides can be analyzed for at least six months if stored in the dark at -20°C (±5°C).

Additional Procedural Recommendations

- The use of a calibrated thermometer is strongly recommended for measuring temperatures of solutions, water baths, and incubators, as these temperatures are critical for optimum product performance.
- Carefully check the temperature of preheated solutions.
- Carefully check the pH value of all solutions. It must be in the range of 7.0 - 7.5 at room temperature.
- The wash concentrations (stringency), pH and temperature are important, as low stringency can result in non-specific binding of the probe and too high stringency can result in lack of signals.
- **Before opening:** Spin briefly to collect probe mix at the bottom of the tube.

FISH Protocol for MetaSystems' DNA Probes

Slide Preparation

1. Spot cell sample onto cleaned microscope slide. Allow to air dry. If you are not using these slides the same day, store at -20°C (±5°C).
2. Apply 10 µl of probe mixture.
3. Cover with coverslip 22 x 22 mm².
4. Seal with rubber cement.

Denaturation

1. Denature sample and probe simultaneously by heating slide on a hotplate at 75°C (±1°C) **for 2 min.**
For tissue sections different denaturation times and temperatures may apply. Please refer to the MetaSystems Tissue Pretreatment kit, D-0905-025-TF.

Hybridization

1. Incubate in a humidified chamber at 37°C (±1°C) overnight.

Post-Hybridization Washes

Solutions Required

- 0.4 x SSC (pH 7.0 – 7.5) at 72°C (±1°C)
- 2 x SSC, 0.05% Tween-20 (pH 7.0) at room temperature

Procedure

1. Remove coverslip and all traces of glue carefully.
2. Wash slide in 0.4 x SSC (pH 7.0) at 72°C (±1°C) **for 2 min.**
3. Drain slide and wash in 2 x SSC, 0.05% Tween-20 (pH 7.0) at room temperature **for 30 seconds.**
4. Rinse briefly in distilled water to avoid crystal formation and let air dry.

Counterstain

Solutions required:


- DAPI/antifade (e.g. MetaSystems DAPI/antifade, D-0902-500-DA)

Procedure:


1. Apply 10 µl of the DAPI/antifade and overlay with a 24 x 32 mm² coverslip.
2. Allow the penetration of DAPI/antifade **for 10min.**
3. Proceed with microscoping and analysis.
4. Store slides at –20°C (±5°C). Hybridization signals are fine for at least six months.

Expected Results


Normal Cell:
Two green-orange (2GO) fusion signals.



Aberrant Cell (typical results):
One green (1G), one orange (1O), and one green-orange (1GO) fusion signal.



Aberrant Cell (typical results):
One green (1G) and one green-orange (1GO) fusion signal.



Only the most frequent signal constellations are shown, other relevant signal patterns may be observed.